

CADIAX 4

*A Practical Method for
TMJ Registration*


Instructions for Use 



"Gamma"
Medizinisch-wissenschaftliche
Fortbildungs GmbH



CE 0483



CADIAX 4

Instructions for Use

for
CADIAX 4, model A
Firmware version 2.13
and
GAMMA Dental Software
version 8.9

Revision:
2026-08-18

© 2026 "Gamma" Medizinisch-wissenschaftliche Fortbildungs GmbH

Wasserzeile 35
3400 Klosterneuburg
Austria

Phone: +43 2243 34140
Fax: +43 2243 34140 90
E-Mail: office@gammadental.com
Internet: www.gammadental.com



Guarantee and Liability

Information in this publication is subject to change without prior notice. "Gamma" Medizinisch-wissenschaftliche Fortbildungs GmbH makes no warranty of any kind and reserves the express right to change this documentation without notification to any other party.

Despite thorough validation and verification of the product to the highest applicable standards "Gamma" Medizinisch-wissenschaftliche Fortbildungs GmbH cannot entirely rule out system malfunction or user error that could lead to the calculation of incorrect evaluations. Any information provided by the software must therefore be subjected to a plausibility check by the dental professional and must not be the sole basis for conducting diagnostic or therapeutic decisions. Final responsibility for the patient always rests with the attending physician. "Gamma" Medizinisch-wissenschaftliche Fortbildungs GmbH accepts no liability.

© 2026 "Gamma" Medizinisch-wissenschaftliche Fortbildungs GmbH

This document contains information protected by copyright. All rights reserved. This document may not be duplicated, reproduced or translated into other languages, fully or partially, without the express written permission of "Gamma" Medizinisch-wissenschaftliche Fortbildungs GmbH.

®, ™ Trademarks and Brands:

Axiograph and SAM are registered trademarks of the company SAM Präzisionstechnik GmbH, Germany. Microsoft and Windows are trademarks of the company Microsoft Corporation, USA. Artex is trademark of the company AmannGirrbach GmbH, Germany. Denar, Hanau and Whip Mix are trademarks of the company Whip Mix Corporation, USA. KaVo is trademark of the company KaVo Kaltenbach & Voith GmbH & Co, Germany. Panadent is trademark of the company Panadent, USA. GAMMA, GAMMA Dental Software, CADIAX and CADIAS are trademarks of the company "Gamma" Medizinisch-wissenschaftliche Fortbildungs GmbH, Austria. Stratos is trademark of the company Ivoclar Vivadent AG, Liechtenstein.

All of the products and product names mentioned in this handbook are brand names of the respective companies. The absence of the symbols ® or ™ does not imply that the name is an unprotected brand name.

Contents

1	Introduction	9
1.1	Explanation of used symbols on packaging and device	9
1.2	Important safety considerations	11
1.2.1	General safety instructions	11
1.2.2	Special information about medical electrical devices	12
1.2.3	Contraindications - Limitations to application	15
1.2.4	Magnet sensor system.....	16
1.3	Intended purpose	16
1.4	About this manual	16
1.5	Symbols and notations.....	16
1.6	User interface terminology	17
1.7	Scope of delivery	18
1.8	System requirements	18
2	Preparing to operate the CADIAX device	19
2.1	Installing the software	19
2.2	Starting and activating the software.....	19
2.3	Setting the face bow	19
2.4	Hardware test.....	20
3	Recording TMJ movement.....	21
3.1	Posture of the patient.....	21
3.2	Instructing the patient.....	21
3.3	Preparations.....	21
3.3.1	Fixing the clutch onto the lower jaw teeth.....	21
3.4	Mounting the kinematic facebow	26
3.4.1	Condylgraph.....	26
3.4.2	Condylgraph comfort.....	28
3.4.3	Preparing the lower face bow	31
3.4.4	Mounting the lower face bow	31
3.4.5	Determining the joint axis	32
3.4.6	Attaching the sensors	33
3.5	Setting up the CADIAX system.....	33
3.6	Software-assisted axis localization	34
3.6.1	Single axis localization.....	34
3.6.2	Dynamic axis localization.....	35
3.6.3	Axis fixup.....	36
3.7	Setting the reference position	36
3.8	Face bow parameter input	37
3.8.1	Condylgraph.....	37
3.8.2	Condylgraph comfort.....	37
3.9	Recording joint movements	38
3.9.1	Joint track curves	38
3.9.2	Condyle position measurement (CPM).....	40

3.9.3	Using recording macros	42
3.10	Finishing the registration	42
3.11	Real-time display	42
3.11.1	Flags.....	43
3.11.2	Time curves.....	43
3.11.3	3D Animation.....	43
4	Mounting the articulator.....	44
5	Troubleshooting	45
6	Background information	48
6.1	The reference coordinate system.....	48
6.2	Exact and anatomic hinge axis	49
6.3	Immediate side shift (ISS)	50
7	Specifications	51
7.1	Technical data.....	51
7.2	Cleaning	51
7.2.1	CADIAX device.....	51
7.2.2	Measuring flags	51
7.2.3	Measuring styli	51
7.2.4	Face bow.....	51
7.2.5	Functional occlusal clutch	52
7.2.6	Occlusal tray clutch	52
7.2.7	Bite fork	52
7.2.8	3D joint support	52
7.3	Recurrent tests.....	52
7.4	Operating conditions	52
7.5	Storage and transport	53
7.6	Disposal.....	53

1 Introduction

Thank you for showing your confidence in us by purchasing the CADIAX system. You have purchased a top-quality medical device, designed based on more than 25 years of experience in the electronic condylography registration.

In the following pages, we would like to help you get acquainted with the CADIAX device and its software.

1.1 Explanation of used symbols on packaging and device

The following symbols are used on the product and/or within the operating instructions:



Attention: See accompanying instructions



Attention: Magnetic field



Follow operating instructions



Medical device



Unique Device Identifier (UDI)



Manufacturer



Manufacturing date




Serial number



Upper and lower limit for temperature



Upper and lower limit for relative humidity



Upper and lower limit for atmospheric pressure



Do not sit on the device



Do not step on the device



Type B applied part



Dispose separately from other waste material



Direct Current (DC)



Universal Serial Bus (USB)



Foot switch

right



Measuring sensors of the right patient side

or



or


Measuring sensors of the left patient side



Phone number



Fax number



E-Mail address



Internet address

1.2 Important safety considerations



For your own safety, as well as for operational safety, please read these instructions thoroughly, before beginning to operate the device. Always comply with any and all warnings included in these instructions, and/or on the device itself.

1.2.1 General safety instructions



The device must be operated in a roomy area, in order to allow for proper dissipation of heat.

Never operate the device in areas where there is a danger of water or other liquids seeping into the device.

The device is powered by the computer it is connected to, through the Universal Serial Bus (USB). Only connect the device directly to a computer's USB terminal. Never operate the device through a USB hub and never use alternative power sources of any kind.

Never set up the system in the proximity of potential fields of electromagnetic interference (loudspeakers, mobile phones, sterilizers, etc.).

Never attempt to repair the device yourself. All repair work must be done by authorized, trained technicians.

In the following cases disconnect the device from the computer and contact a qualified service technician:

- If water or other liquid seeps into the device.
- If the device does not operate properly, although the operating instructions have been followed correctly.
- If the device has fallen down or if the housing is damaged.
- If the device shows noticeable deviation from its normal operation.

Please make sure that the device is set up on a smooth, even surface, which is longer and wider than the device itself.

The surface on which the device stands must be sufficiently stable, as the device could be badly damaged by shaking or falling.

In choosing the working area and/or storage area, make sure that the device is not subject to extreme deviations in temperature or humidity, as well as to direct sunlight and excessive heat.

Make sure that the device is not subjected to vibrations and jolts.

Also make sure that the cable does not pose a hindrance (tripping hazard).

Do not sit on the device since this may cause it to be severely damaged.



Do not step on the device since this may cause it to be severely damaged.

1.2.2 Special information about medical electrical devices


1.2.2.1 Setting up the CADIAX device



If the computer connected with the CADIAX device is in the proximity of the patient, it must have been tested in compliance with standard EN 60601-1.


The patient proximity (definition according to standard EN 60601-1) is the area within which the patient could, intentionally or unintentionally, come into contact with:

- either any parts of the medical electrical device/system, or
- with someone using or touching a part of the device/system.




Definition of patient proximity, according to standard EN 60601-1

If the computer connected with the CADIAX device is not built according to standard EN 60601-1, it must be set up outside the proximity of the patient.



If such a computer is set up inside the patient's proximity, it and all devices connected to it (e.g. monitor) have to be isolated according to EN 60601-1. Power sources have to be connected through a medical isolation transformer and tethered network connections through a medical network isolator.



Avoid using the CADIAX device directly next to other devices or stacked on top of other devices, as this could lead to faulty operation. If, however, it is necessary to use the device in the manner stated, the CADIAX device as well as the other devices must be monitored to ensure they are working properly.

1.2.2.2 Operating the CADIAX device



Only switch the device into recording mode after the measuring sensors (flags and styli) are completely mounted.


Make sure that the sensors are not touched during the registration, not by yourself nor by the patient.

Always switch off the recording mode before detaching the flags and styli from the face bow.

1.2.2.3 Features of the CADIAX 4 device enclosure (diagnostic sensor set)


When using the "diagnostic" sensor set, the enclosure of the CADIAX 4 main unit has the following features:

- 1 Connector socket for the measuring stylus of the right patient side
- 2 Connector socket for the measuring flag of the right patient side
- 3 LED for indication of the current device status (refer to chapter "Status indication of the CADIAX device")



Front side

- 4 Connector socket for the measuring flag of the left patient side
- 5 Connector socket for the measuring stylus of the left patient side




Back side

- 6 Connector socket for the foot switch
- 7 Type label with information regarding the manufacturer and the device
- 8 USB 2.0 (type B) connector socket for the cable connection to the PC


1.2.2.4 Status indication of the CADIAX device

The CADIAX device indicates its operating state via a green LED on the front side of the enclosure. This LED can be in the following conditions:



LED is permanently on:

Device is connected to a PC and is ready for operation.



LED flashes with 5 Hz:

A measurement is in progress.



LED flashes with 0.5 Hz:

An internal hardware error occurred in the device. No measurement is possible.



LED permanently off:

The device is not connected to a PC and is turned off.

1.2.2.5 Electromagnetic compatibility



Medical electrical devices require special safety precautions with regard to electromagnetic tolerance. Therefore, the CADIAX device may only be installed and put into operation according to the instructions contained in this manual. Failure to comply with the instructions herein can potentially impair the measurement accuracy stated in chapter "Technical data".



Only use accessories which are explicitly approved for CADIAX devices. Using other accessories, especially cables which are longer than the cable lengths prescribed below, can lead to increased emission of interferences, or to reduced interference immunity of the CADIAX device, which may lead to faulty operation.

The following accessories to the CADIAX system affect the system's electromagnetic compatibility and may be replaced by the user. Components not listed here may only be replaced by the manufacturer.

Accessory	Maximum cable length
Foot switch	3.0 m
USB cable (USB 2.0, type A plug to type B plug, shielded)	3.0 m
CADIAX "compact" sensors	1.5 m
CADIAX "diagnostic" stylus (only applicable to CADIAX 4 devices)	1.5 m
CADIAX "diagnostic" flag (only applicable to CADIAX 4 devices)	1.5 m



Portable and mobile HF-communication appliances, e.g., mobile telephones, cordless telephones, computers, and laptops with activated WiFi, etc., can have detrimental effects on medical electrical devices. Do not operate the CADIAX device in the proximity of these potential sources of interference, and make sure to adhere to a safe distance of at least 30 cm. Non-compliance can lead to reduced performance of the device.

1.2.2.6 Connecting other devices



Devices connected to analog or digital interfaces must be certified to satisfy the applicable standard EN specifications (e.g., EN 60950 for data-processing devices, or EN 60601 for medical electrical devices). In addition, all configurations must satisfy system standard EN 60601-1. Whoever connects supplementary devices to the signal input or outlet units is the system configurator and therefore responsible for ensuring that system standard EN 60601-1 is adhered to.

1.2.3 Contraindications - Limitations to application



When operating the CADIAX system, the following conditions can hinder or limit its application:

- **Cognitive abilities:** Patients who are not able to follow or carry out the attending dentist's instructions correctly (e.g., where to move the lower jaw, etc.), for physical or psychological reasons.
- **General clinical symptoms:** Patients with illnesses that do not allow for the attachment of a face bow for registration purposes. Illnesses may be physical or psychological in nature (e.g., spastic, epilepsy, claustrophobia, injury or disease of the skull or soft-tissue structures of the skull, ears, etc.).
- **Dental clinical symptoms:** Patients with odontopathy or periodontal disease, which excludes the mounting of a clutch (parodontosis, loose teeth, damage to the tooth substance, changes in the mucous membranes in the mouth or pharynx, etc.).

- **Certain diseases in the ear region:** Patients with diseases in the ear region which exclude the fixation of the face bow in the auditory canal (e.g., painful inflammations, tinnitus, ear pain, etc.).
- **Muscle pain in the head and neck region:** Patients with extreme muscle pain in the head and neck region that are not able to wear the face bow for the duration of the examination.
- **Patients between 0 and 9 years of age:** Experience has shown that children are first able to follow the dentist's instructions well enough to lead to a successful examination from about 10 years of age onwards. In addition, existent deciduous teeth can make it difficult to fix the clutch.
- **Wearers of implants:** Patients with active implants (e.g., cardiac pacemakers) should be handled cautiously when using the CADIAX magnet-sensor system, as the stylus tip, to which a magnet is attached, could come into proximity of such implants.

1.2.4 Magnet sensor system



The CADIAX Magnet-Sensor system ("M"-Sensor) contains a magnet in the tip of the stylus. For this reason, special care is necessary when using it with patients with cardiac pacemakers or implanted defibrillators. Make absolutely sure to adhere to the necessary safe distances! If necessary, consult the responsible cardiologist. If in doubt, refrain from using the CADIAX "M" Sensor system.

1.3 Intended purpose

The CADIAX system is designed for registration and display of hinge axis movements of the human mandible, conversion of these movements to the intercondylar distance of an articulator, and calculating the settings for this articulator, appropriate to the patient.

The system is used by dentists and dental technicians.

1.4 About this manual

The authors of this user manual assume that the reader has a basic knowledge of condylography, working with a face bow, and the transfer of its head-related data into an articulator.

In addition, a sound basis in the Microsoft Windows operational system is required. The user should be able to work with the user-interfaces and other operational elements of the Windows programs.

This handbook is not a substitute for a basic education in using the Windows operating system. If you have only little or no education in Windows, we recommend that you first become acquainted with the system. This will make it much easier to work with the CADIAX software.

1.5 Symbols and notations

Special attention will be brought to important aspects using the following symbols:

	Warning	Possible sources of error, or the danger of a possible operational error.
	Attention	Especially important information regarding the topic being described.
	Information	Useful tips and tricks.

The following symbols and notations will be used in this user manual to describe certain commands and instructions for operating the software:


- Designations of keys will be written in italic font style. For combinations of keys, the individual key specifiers will be concatenated with plus signs (+). For example, *Ctrl+D* means: Hold down the *Ctrl* key on your keyboard and simultaneously press the *D* key.
- Names of menu items in the software will be written in italic font style. If several menu items need to be selected in succession, they will be concatenated with arrows. For example, *File → Close* means: First, click on the *File* menu item in the menu bar, then click on the menu item *Close* in the opened menu.

The first occurrence of ***Special Terms*** that may be used in subsequent descriptions are highlighted in bold and italic font style.


1.6 User interface terminology

All GAMMA software applications use a user interface layout consisting of the following parts:


- The ***Working Area*** is the area of the application that displays the actual patient data and will therefore receive the highest amount of attention and interaction.




- The hierarchical menu structure of the ***Menu Bar*** at the upper border of the window provides access to many of the application's functionalities. Most general application settings that are independent of the actual patient data can be configured here.



- The ***Tool Bar*** that is located below the menu bar provides icons to quickly access commonly used functionalities. The availability of these icons can depend on the actual state and working context of the application.



- The **Status Bar** at the lower border of the window displays contextual information regarding the actual state of the application as well as progress information in the case of long-running tasks.



1.7 Scope of delivery



The accompanying packing list describes the exact scope of delivery. After taking delivery of your CADIAX system, please make sure to immediately check the package contents for completeness and faultless condition.

The CADIAX system does not include the material for fixing the occlusal-tray clutch on the lower jaw teeth of the patient. You will need suitable bite registration material (e.g. Exabite II by GC) and the appropriate dispenser.

For fixing the para-occlusal clutch, you will need a composite for temporary restorations (e.g. Protemp II by 3M) and a temporary cement such as a zinc-oxide cement (e.g. TempBond by Kerr) or a carboxylate cement (e.g. Durelon by 3M).

Furthermore, a face-bow system is required for mounting the CADIAX measuring sensors, such as Condylgraph or Condylgraph comfort by GAMMA. Such is not included in the regular scope of delivery of the CADIAX system.



Please contact your local distributor for further information.

1.8 System requirements

A computer with installed GAMMA Dental Software is required in order to use the CADIAX system.



Please refer to the manual of GAMMA Dental Software for the requirements regarding the computer system.

2 Preparing to operate the CADIAX device

Before beginning to work with the CADIAX system, several preparatory steps must be carried out.

2.1 Installing the software



Install the software before connecting the CADIAX device to your computer for the first time. This will install the necessary drivers to let your operating system to recognize the device.




Refer to the manual of GAMMA Dental Software for the installation of the software.

2.2 Starting and activating the software




The procedure of starting and activating the software is described in the manual of GAMMA Dental Software.

Following a successful activation, the recorder application CADIAX Recorder can be started from GDSW classic or GAMMA Document Browser and stores the recorded data in the respective database.




2.3 Setting the face bow

Set the type of the face bow that you use in the menu *Options* → *CADIAX Unit Settings*:



The various face bows have different scales for entering face geometries. If a wrong face bow is selected or a wrong scale value is entered, the recordings can not be recalculated correctly. This can lead to a distorted picture of the recorded condylar tracks and, in some circumstances, to a falsification of the calculated value settings for the articulator.

The face bow *Condylograph comfort* may be used with either of the following reference point indicators.



The fixed reference point indicators P (Patient) and T (Transfer).




The individual reference point indicator.




The fixed reference point indicators are mounted at a height of 22 mm, which cannot be adjusted. If the anatomical situation of the patient requires a different height, the individual reference point indicator should be used.

2.4 Hardware test

The pane *Hardware Test* of the CADIAX Recorder software allows checking the hardware condition of the system.



The display indicates errors in the system by highlighting the affected component in the graphic and shows an error description in the area *Diagnostics*.



3 Recording TMJ movement

The registration of mandibular movement is always carried out in the following steps:

1. Preparations
2. Fixing the clutch onto the teeth of the lower jaw
3. Mounting the upper face bow
 - a. Condylograph
 - b. Condylograph comfort
4. Mounting the writing bow
5. Locating the hinge axis
6. Attaching the measurement flags and styli
7. Recording the joint track curves
8. Registration of CPM (if a para-occlusal clutch is used)
9. Finishing the registration, saving data

3.1 Posture of the patient

The patient should sit in an upright and relaxed position. Body movements during condylography recording can change the hinge axis position and should be avoided. The head should be held upright and the upper spine in the habitual normal position, i.e. not be bended forward or backward. It is recommended to use a head and neck support.

Glasses, earrings, hair barrettes, etc. of the patient should be taken off. Electrostatically charged hair can cause artifacts in the recordings. This effect is intensified by hairspray. For this reason, the patient should wear a cap, e.g. a surgical cap as used by surgeons. Long hair should be combed back behind the ears.

3.2 Instructing the patient

Explain the lower jaw movements like protrusion, retrusion, mediotrusion and mediotrusion to the patient. Ask the patient to execute all these movements completely and as far as possible. Practice these movements before mounting the face bow.

3.3 Preparations

Before starting the registration, prepare all necessary devices and materials.

Connect the CADIAX device to your computer and start the recording software. If the device has been detected, this will be indicated in the status bar of the software. If this is not the case, you can find information for error correction in the chapter "Troubleshooting".

3.3.1 Fixing the clutch onto the lower jaw teeth

There are two different kinds of clutches available for attaching the lower face bow to the lower jaw. These are:



The occlusal-covering tray clutch.

The para-occlusal clutch.

In the case of edentulous patients, the clutch can also be attached to the lower jaw using a mandibular clamp.



The occlusal-covering tray clutch and the functional occlusal clutch are shipped in non-sterile condition and must be disinfected and/or sterilized before use. For the hygienic instructions regarding cleaning and reuse, please refer to chapters "Functional occlusal clutch" and "Occlusal tray clutch", respectively.



Examine the periodontal status of the lower jaw dentition. In case of a poor periodontal situation, either refrain from using the clutch, or, at the attending dentist's responsibility, take suitable precautionary measures (e.g., blocking out). Areas of the dentition, undercut to the occlusal plane, should always be blocked out with appropriate material (e.g., soft wax), even with patients that have a healthy periodontal status.

If you fix the clutch to a provisionally cemented denture, the possibility arises that the denture can become loose due to mechanical influences. In such cases, you must determine whether provisional cementing is appropriate or not.




Check whether the clamp on the writing bow can be pushed onto the clutch shaft easily. If this is not the case, do not continue using that particular clutch.

3.3.1.1 Fixing the occlusal tray clutch



If you are using a para-occlusal clutch, you may skip this section.

The occlusal tray clutch is available in multiple sizes. Select a clutch of an appropriate size for the patient.



Before introducing the silicone material, inlay the clutch with 3 thin rolls of soft wax (diameter 3-4 mm). The rolls are attached at diagonals to the arch, one in the front and two in the back. They provide predetermined breaking points, which will make removing the clutch much easier.



Use exclusively a silicone adhesive (bite registration material), with a Shore-hardness of maximum A60. Follow the manufacturer's instructions!

Fill the clutch to maximum 2/3 with the silicone material.



Place the filled clutch onto the lower jaw teeth and set the shaft in a median-sagittal position. The patient draws back the lower jaw and closes, thus automatically positioning the shaft so that the smallest possible locked bite is achieved.



Before mounting the lower face bow, the silicone must be completely set, with the patient's mouth slightly open. Follow the manufacturer's instructions! It is advisable to leave a separate material sample extra-orally with which you can check the hardening process. Make sure that no interferences occur in the back areas, caused by material being squeezed out.

3.3.1.2 Fixing the para-occlusal clutch



If you are using an occlusal tray clutch, you may skip this section.

When using a functional occlusal clutch, you can analyze the effects of occlusion on the TMJ position and hinge axis movement. Condyle position measurements (CPM) can be carried out directly in the patient's mouth, so you no longer need to measure CPM on mounted casts. The functional occlusal clutch is located outside of the occlusion, so it doesn't touch the upper jaw teeth, neither in intercuspation, nor in function.


For fixing the para-occlusal clutch you will need a composite for temporary restorations (e.g. Protemp II by 3M ESPE).



Additionally, you will need a temporary cement such as a zinc oxide cement (e.g. TempBond by Kerr) or a carboxylate cement (e.g. Durelon by 3M ESPE).



Place the clutch in the patient's mouth and bend the wings to fit on the lower tooth arch. Allow 1-2 mm of space between the teeth and the clutch, for the plastic.



Place a wax plate between the upper and lower teeth, and have the patient bite down. The wax plate prevents the plastic from flowing onto the upper teeth, and eliminates the tiresome job of having to polish the clutch's plastic surface.



Prepare the composite according to the manufacturer's information.



Coat the surface of the clutch, which should be facing the labial and buccal teeth, with a 2 mm thick layer of plastic. Wait until the material begins to set.



Place the clutch in the patient's mouth and instruct the patient to close, without allowing the upper teeth to touch the clutch (check this).

The clutch shaft should lie in the median sagittal plane.

Wait until the material has set, then remove the clutch from the patient's mouth.



Remove any excess interferences of the plastic material with a sharp knife or a cutter.



Use as little plastic as possible. Too much plastic often causes pain in the gums.

Check to see whether the clutch fits properly in the mouth.



Occlusion paper marks the contacts between the clutch and upper jaw teeth, which have to be removed.



If the plastic has not completely set yet, you can lay the clutch aside and begin mounting the upper face bow.

Prepare the cement according to the manufacturer's information.

Fix the clutch with the cement on the lower jaw teeth.





Do not fix the para-occlusal clutch onto painful or loose teeth.

After the registration, the clutch can be removed easily. The teeth will only need a light cleaning (scaling). The same plastic-coated clutch can be re-used for the same patient (residues on the clutch do not necessarily need to be removed).

3.4 Mounting the kinematic facebow

3.4.1 Condylograph

3.4.1.1 Preparing the upper face bow

Before mounting the Condylograph on the patient, the forehead bracings of the upper face bow need to be pulled back.

Slightly loosen the side arm clamp screws, so that the side arms are movable in sagittal (upper clamp screw) and transversal (lower clamp screw) direction.

Loosen and remove the orbital pointer.



The upper face bow is now ready to be mounted.

3.4.1.2 Mounting the upper face bow

Adjust the width of the face bow in a way, that the side arms are in contact with the skull above the ears evenly without tension. The face bow should be adjusted symmetrically.



Put the face bow on the patient's head, whereby the glabella support is placed on the glabella and the side arms rest above the ears.



Lower the forehead supports until they are in skin contact and fixate them. An even support on forehead and glabella should be achieved, which stabilizes the face bow and significantly improves the patient's comfort.



The retention straps are placed parallel to the side arms and fastened. Now, the face bow should hold firmly on the head without further support.



Assemble the red writing flags with writing paper and fixate them on the face bow.



Adjust the depth of the side arms so that the posterior edge of the flag covers the tragus.



Pay attention that the side arms lie above the ears. The face bow must not be pulled downwards, onto the pinna, or upwards by the retention straps.

Attach the reference point indicator on the cross bar of the face bow and adjust it to the orbital.



To indicate the position of the orbital point with a lead mark in a subsequent x-ray, you can now mark the point on the skin.

Afterwards, fixate the reference point indicator and remove it from the face bow.



3.4.2 Condylograph comfort

3.4.2.1 Preparing the upper face bow

Slightly loosen the side arm clamp screws, so that the side arms are movable in sagittal (upper clamp screw) and transversal (lower clamp screw) direction.

Loosen and remove the orbital pointer.



Fixate the glabella support by pushing it from the side into the center position until it snaps in.



The actual nose support can easily be exchanged. It is only put on the glabella bridge.



The upper face bow is now ready to be mounted.

3.4.2.2 Mounting the upper face bow

Adjust the width of the face bow in a way, that the side arms are in contact with the skull above the ears evenly without tension. The face bow should be adjusted symmetrically.



Put the face bow on the patient's head, whereby the glabella support is placed on the glabella and the side arms rest above the ears.



The retention straps are placed parallel to the side arms and fastened. Now, the face bow should hold firmly on the head without further support.



Assemble the red writing flags with writing paper and fixate them on the face bow.





Adjust the depth of the side arms so that the posterior edge of the flag covers the tragus.



Pay attention that the side arms lie above the ears. The face bow must not be pulled downwards, on the pinna, or upwards by the restraining band.

Attach the reference point indicator on the cross bar of the face bow and adjust it to the orbital.



To indicate the position of the orbital point with a lead mark in a subsequent x-ray, you can now mark the point on the skin.

Afterwards, fixate the reference point indicator and remove it from the face bow.



3.4.3 Preparing the lower face bow


Move the side arms of the lower face bow into their center position.



Slightly loosen the screws, so that the side arms and the double clamp on the crossbar are movable.



Adjust the two axis needles so that they protrude 2 cm from the axis pin jack.



The writing bow is now ready to be mounted.

3.4.4 Mounting the lower face bow

The side arms of the lower face bow are pushed completely outwards, then slightly fixated. The patient holds the lower jaw closed and in centric (retral) position. The double clamp is pushed onto the clutch.

Make sure not to loosen the clutch and check once more that it fits correctly.



Adjust the writing bow so that it is parallel to the upper face bow.

Tighten the clamp on the lower cross bar. In doing so, support the lower face bow with your second hand to avoid too much force being transferred to the clutch. Strong force can cause pain or loosen the clutch.



Insert the axis needles into the retainers at the end of the side arms and adjust their location to the approximate position of the joint axis.

Finally, tighten the screws to fasten the clamps at the lower ends of the side arms to the cross bar.



3.4.5 Determining the joint axis

Push the axis needles in until they are in contact with the writing flag. Fixate the needles in this position.



Start in retral joint position and guide the lower jaw of the patient with the chin in retral open-close movements.



Adjust the setscrews of the side arm in a way that the axis needle solely rotates without drawing a translatory movement on the flag.




3.4.6 Attaching the sensors



Only touch the flags on their borders. Avoid touching the black measuring surface to prevent it from being soiled or scratched. Also make sure that recording mode is turned off while you mount the flags and styli.

Fixate the flags on the designated sockets of the upper face bow. Avoid jamming anything between the side arms and the flags in this process.



The double-stylus system allows for a recording of the hinge axis rotation on both sides.

Insert the stylus pin with the larger case completely into the hole of the side arm. The second pin should be positioned above the first one. Tighten the screw at the end of the side arm to fixate the stylus.

If the system is mounted properly and the lower jaw is in centric position, the stylus tip should be located in the area of the upper back quarter of the flag.



The styli should be positioned in a way that their axis protrude by about 1 cm to the outside. This range of movement is, among other things, required for mediotrusive movements.



Avoid touching the measuring surface of the flags and the stylus pins during the measurement procedure and preparatory work. This also applies to the patient.

3.5 Setting up the CADIAX system

Plug the connector of the foot switch into the designated jack at the backside of the device.



Connect the device to a free USB-port of your PC.



Subsequently, connect the sensors to the device according to the labeling on the casing.



3.6 Software-assisted axis localization


The software provides both single and dynamic axis localization methods.



Basis for recording TMJ movements always has to be a mechanical hinge axis localization (refer to chapter "Determining the joint axis"). The software-assisted hinge axis localization may be used to check the correct adjustment of the lower face bow before or during the recording.

3.6.1 Single axis localization

During single axis localization, the system takes two measurements; one with the patient's mouth being open and one with it being closed. From these two measurements, the software is able to calculate the point around which the styli were rotated and therefore the location of the axis.




1. Explain the procedure to the patient.
2. Guide the lower jaw of the patient into the reference position. Try to avoid tooth contact so that you can position the lower jaw in joint relation without occlusal interference.
3. Press the foot switch to measure the first position. This will initiate a countdown for the measurement of the second position, indicated by a progress bar in the appearing dialog window.
4. Ask the patient to open their mouth about 15 mm. Make sure that the jaw remains in a retral position and does not carry out any translatory movements. Wait for the system to take the second measurement.
5. Let the patient move into reference position and continue with adjusting the styli to the determined axis location (refer to chapter "Axis fixup").



It is often easier to let the patient open their mouth for the first measurement and then move into a retral, centric position. Since that position is also used for fixing the axis, an additional source of error can be avoided.

3.6.2 Dynamic axis localization

Compared to single axis localization, dynamic axis localization uses continuous open/close movements for determination of the joint axis. Multiple measurements are taken and averaged for a more precise result, even with only minor movements.




If the "Rotate Stylus" symbol appears during the axis localization, it indicates that a stylus is in an inappropriate position in relation to the rotation axis of the mandible. You can resolve this by rotating the stylus of the respective side in its retainer.




If the "No Contact" symbol appears during the axis location, it indicates that a stylus has lost contact with the flag surface. If all cables are connected correctly and the stylus is within the sensitive area of the flag, please inspect the flag surface and clean it if necessary.

1. Explain the procedure to the patient.
2. Guide the lower jaw of the patient into the reference position. Try to avoid tooth contact so that you can position the lower jaw in joint relation without occlusal interference.
3. Press the foot switch to start the measurements.
4. Guide the patient's lower jaw in repeated open/close movements. Make sure that the jaw remains in a retral position and does not carry out any translatory movements. The software will display the calculated axis points and the number of measurements taken in real-time.



5. After enough measurements have been performed, the software will indicate this by displaying the calculated axis point in green color on both sides. As soon as that is the case, you can press the foot switch again to stop the measurements.



6. Let the patient move into reference position and continue with adjusting the styli to the determined axis location (refer to chapter "Axis fixup").




The color of the calculated axis point provides information about the number of individual measurements. With up to five measurements, it will be red, with up to ten measurements, it will be blue, and from ten measurements on, it will be green.




You can find advanced options for the calculation of the axis points in the menu *Options* → *Axis Location Parameters*.

3.6.3 Axis fixup

When the axis points have been localized, the styli need to be adjusted to these positions. During the axis fixup, the calculated axis points are displayed as white circles and the actual stylus positions as red crosses.



1. Make sure that the patient is in reference position and stays in that position during the axis fixup.
2. Adjust the inclination and length of the lower bow's side arms so that the red crosses are within the white circles.



3. As soon as the styli are positioned close enough to the axis points, you can press the foot switch to continue with setting the reference position and entering the face bow parameters.




During axis fixup, the distance between stylus position and axis point is displayed at the lower border of the window in millimeters. As soon as this distance falls below a certain threshold, the view will be enlarged automatically to allow more accurate adjustments.



You can mark positions, show movement tracks and change zoom settings via the *View* menu.

3.7 Setting the reference position

Guide the patient into the retral reference position and press the foot switch. The position set here will be used as zero point for the following recordings.



You can also re-set the reference position between recordings via the "refix" functionality or by switching to this tab.

3.8 Face bow parameter input


In order to recalculate the recordings according to the mounting of the upper face bow, it is necessary to enter the face bow parameters.

 Refer to chapter "Setting the face bow" for instructions regarding the selection of the used face bow.

 Note that the terms "left" and "right" always refer to the corresponding side from the patient's point of view. All values have to be entered in millimeters.

3.8.1 Condylograph

The following parameters have to be entered when using the Condylograph face bow.



Y-Shift is the shift of either side arm of the upper bow to left or right side. It can be read from the scales on the front of the face bow's cross bar. Values that are located towards the inner side, relative to the zero point of the scale, are entered as negative values.


X-Shift is the shift of either side arm of the upper bow in anterior or posterior direction. It can be read from the scale on the respective side arm. Values that are located towards posterior, relative to the zero point of the scale, are entered as negative values.

The **anterior face bow height** is the difference in height between the marked orbital point and the upper face bow. The value can be read from the vertical scale of the reference point indicator.

The **anterior face bow depth** is the difference in depth between the marked orbital point and the upper face bow. The value can be read from the horizontal scale of the reference point indicator.

3.8.2 Condylograph comfort

The following parameters have to be entered when using the Condylograph comfort face bow.



Y-Shift is the shift of either side arm of the upper bow to left or right side. It can be read from the scales on the top side of the face bow's cross bar.

X-Shift is the shift of either side arm of the upper bow in anterior or posterior direction. It can be read from the scale on the respective side arm.

The **anterior face bow height** is the difference in height between the marked orbital point and the upper face bow. The value can be read from the vertical scale of the reference point indicator.

When using the fixed reference point indicators, the value cannot be changed and is fixed to their mounting height of 22 mm.


The **anterior face bow depth** is the difference in depth between the marked orbital point and the upper face bow. The value can be read from the horizontal scale of the reference point indicator.

When using the individual reference point indicator, it is additionally required to specify the position, in which this value is measured. In **position T (transfer)**, the pointer is oriented parallel to the side arms of the face bow, and in **position P (patient)**, it is inclined by 30° towards the patient's nose.

3.9 Recording joint movements


3.9.1 Joint track curves

After setting the reference position and entering the face bow parameters, you can start with the actual recording of condylography curves. To do so, change to the *Recording* tab.



If you frequently use the same sequence of recordings, you can create a reusable macro for this purpose. See chapter "Using recording macros" for further information.

CADIAX Recorder allows an unlimited number of recordings to be performed. To create a new curve recording, select the menu point *New Curve* in the *Data* menu or the corresponding button in the toolbar. The dialog for specifying the parameters of the recording will appear.



Type in this dialog refers to the name of the curve that will be passed on to the analysis applications and remains editable in those.

Time specifies how long the movements of the patient will be recorded. A duration of 4.5 seconds turned out to be reasonable for this purpose. For movements that take more time (chewing, speech, etc.), it is possible to increase the duration accordingly.




Before starting a recording, ensure that the patient is in reference position. In the ideal case, the red cross of the real-time cursor is positioned exactly in the origin of the coordinate system. If this position is difficult to reproduce, you can re-set the software's reference position via the tab of the same name (refer to chapter "Setting the reference position").




The provided system- and user-defined recording types can be configured via the menu *Edit* → *Definitions*. Instead of selecting a provided type, you can also enter a custom name for the recording.


After having confirmed the recording parameters by clicking **OK**, you can start the recording by pressing the foot switch. Subsequently, have the patient carry out the selected movement.



The start and the end of the recording are signaled acoustically. During the recording, the progress will be displayed and the curve will be drawn in real-time.



Once the recording is finished, it will be displayed in the recording window and you can proceed with the next one.



You can use the menu item **Data → Delete** or the corresponding toolbar button to remove an already performed recording.



Check a curve for its quality immediately after the recording. To do so, check the starting point of the movement track relative to the reference position (i.e. the origin of the coordinate system), as well as smoothness of the curve line. Ideally, the movement starts exactly in the zero point and doesn't show any distortion artifacts of any kind.

Registrations with obvious distortions should not be accepted. In such cases, clean the flags and styli thoroughly and repeat the recording. If this doesn't prevent artifacts from appearing, cancel the registration and contact an authorized GAMMA service center.

In addition, movement tracks that differ from the expected shape should be treated skeptically and have to be verified with further registrations.



Please note that the designations mediotrusion right and mediotrusion left always refer to the condyle of the patient that is carrying out the mediotrusive movement. Mediotrusion right therefore means a mandibular (i.e., of the chin) movement to the left side and vice versa.



To overlay a curve with one another, you can have the former permanently displayed in the background by using the menu item **Data → Selected As Reference Curve**, the corresponding toolbar button, or by simply double-clicking the item in the list.


3.9.2 Condyle position measurement (CPM)

A condyle position measurement (CPM) describes the measurement of the 3-dimensional offset in the joint between a reference position and any another joint position (e.g. maximum intercuspation ICP). For CPM, the first measurement of the recording will be the reference to the following ones.



Naturally, measurements of a position relative to the occlusion (e.g. ICP) only make sense if the registration clutch is fixed to the lower jaw teeth in para-occlusal method.

To create a new CPM recording, select the menu point *New CPM* in the *Data* menu, or the corresponding button in the toolbar. The dialog for specifying the parameters of the CPM recording will appear.



Similar to the curve recording parameters, **Type** is the name assigned to the CPM. A recording of type *Point Mode* can hold an unlimited number of individual point measurements, which are taken by pressing the foot switch repeatedly.

For CPM, **Time** refers to the amount of time between the first and the second measurement. *Impulse* stands for an unspecified amount, where the second measurement will be taken as soon as the foot switch is pressed again.




The provided system- and user-defined recording types can be configured via the menu *Edit* → *Definitions*. Instead of selecting a provided type, you can also enter a custom name for the recording.


The predefined acronyms stand for:

- RP: Reference Position
- RCP: Retral Contact Position
- ICP: Intercuspation
- RES: Resilience of the mandibular joint
- ETP: Expected Therapeutic Position
- IVP: Ideal Vertical Position
- FBP: Forced Bite Position


After having confirmed the recording parameters by clicking *OK*, you can start the recording by pressing the foot switch.



The start and the end of the recording are signaled acoustically. If a time between the two measurements has been specified, the progress will be displayed during the recording.



Once the recording is finished, it will be displayed in the recording window and you can proceed with the next one.



You can use the menu item **Data** → **Delete** or the corresponding toolbar button to remove an already performed recording.


By selecting the CPM points in the coordinate system, you can have the following values displayed on the lower border of the window:

- Delta X** The distance between the two measurements along the X axis, i.e. in sagittal direction (+X = anterior, -X = posterior), for right and left side.
- Delta Y** The distance between the two measurements along the Y axis, i.e. in transversal direction (+Y = to the right side, -Y = to the left side).
- Delta Z** The distance between the two measurements along the Z axis, i.e. in vertical direction (+Z = downwards, -Z = upwards), for right and left side.
- Delta H** The difference of the incisal pin height in the articulator.
- Delta W** The side shift of the incisal pin on the incisal table.
- Delta L** The anterior-posterior shift of the incisal pin on the incisal table.
- Gamma** The amount of hinge axis rotation between the two measurements.

3.9.3 Using recording macros

With recording macros, you can define reusable sequences of curve and CPM recordings once and repeat them during following recording sessions.

The macro editor allows you to create new macros or edit existing ones. It can be started via the item *Edit Macro File* in the *Data* menu.



Use the provided buttons to add curve or CPM recordings to the sequence or to remove them. You can change the order of the items in the list by dragging them while holding the left mouse button. Once finished, save the macro to a file using the corresponding button.

To use a created macro during recording, switch to the Recording tab and select the menu entry *Data* → *CADIAX® Recorder Macro* or the corresponding button in the toolbar.



The option *Macro Easy Start* in the *Options* menu allows individual recordings to be started with only a single press of the foot switch during the execution of a recording macro.

3.10 Finishing the registration

After having performed all required registrations, you can transfer them to GAMMA Dental Software via the menu item *Data* → *Save All* or via the floppy disk icon in the toolbar.



Make sure to switch off the recording mode before starting to dismantle the system!



To prevent data loss due to issues occurring while saving the recordings, those are saved in the sub-directory *GDSW/Temp* of the user's documents folder. To recover such a file with the extension *.cxc, use the corresponding menu item in the *Data* menu.

Disassemble the system in reverse mounting order. First, loosen the measuring styli and flags from the face bow and subsequently remove the lower face bow from the clutch. Afterwards, remove the upper face bow and eventually remove the clutch.




To loosen the clutch, move its shaft carefully up and down. When removing the clutch, take care to not have it hit against the teeth of the upper jaw due to jerky movements. If necessary, protect them, e.g. by putting a wax plate in between the teeth.

3.11 Real-time display

CADIAX Recorder provides several ways to visualize the stylus movements on the measurement flags in real-time, without performing a recording.

3.11.1 Flags


The pane *Flags* displays the physical X/Z positions of the stylus pins on the flags.



You can mark positions, show movement tracks and change zoom settings via the *View* menu.

3.11.2 Time curves

The pane *Time curves* displays the measured coordinate and AUX values over time. The upper display area represents the right patient side, the lower area the left one.




You can adjust the display speed and the curves to display via the area on the right border of the window.



AUX data from external sources can only be processed with CADIAX diagnostic devices.

3.11.3 3D Animation


The pane *3D Animation* visualizes the measured jaw movements with a virtual representation of the patient's skull.




The area on the left side of the window provides miscellaneous display options.

4 Mounting the articulator

The CADIAX system uses the axis-orbital-plane as reference plane for recordings and for transferring the jaw position into the articulator (refer to chapter "The reference coordinate system"). In the articulator, the plane is described by the right and left hinge axis point and the contact plane of the incisal table.



This ensures that, when the upper jaw model is transferred into the articulator using the face bow, the calculated values will match the actual settings of the articulator's condylar housings. For transferring the model position, a bite fork is attached to the face bow via a freely movable 3D joint support. An impression of the upper jaw teeth is taken and replicated with the plaster models in a transfer stand or directly in the articulator.



If a different face bow is used for transfer, make sure that the face bow is set to use the same reference points. Only then is it guaranteed that the values calculated by the CADIAX system match the model situation in the articulator.

For detailed information regarding the face bow transfer and the mounting of models in the articulator, please refer to the instruction manual "Reference SL Articulator and Face Bow Systems".

5 Troubleshooting

There are always two steps involved in correcting errors in the CADIAX system: first, determine the problem and then carry out the recommended solution to eliminate the error.

If errors occur that are not covered in this chapter, contact an authorized GAMMA service center to resolve the problem.

The device is not detected by the recording software and the green LED on the front of the device is unlit.

Possible cause of error	Solution
The device is not connected to the computer.	Connect the device with an USB port of your computer.
The USB connection between device and computer is defective.	Use another USB port or another USB cable.
The CADIAX device is defective.	Contact an authorized GAMMA service center.

The device is not detected by the recording software but the green LED on the front of the device is lit.

Possible cause of error	Solution
The USB connection between device and computer is defective.	Use another USB port or another USB cable.
The CADIAX device is defective.	Contact an authorized GAMMA service center.
The driver software has not been installed correctly or has not been installed at all.	Disconnect the CADIAX device from the computer. Start the setup routine of the software and perform a repair. Restart the computer and reconnect the device after a short waiting period. Start the recording software only when the operating system has finished configuration of the device. If the computer is connected to the internet, you can also have the drivers installed by Windows Update automatically.
Other software-related issue.	Contact an authorized GAMMA service center.

The green LED on the device front blinks slowly and/or the recording software notifies that a hardware defect has been encountered.

Possible cause of error	Solution
The CADIAX device is defective.	Contact an authorized GAMMA service center.

After mounting, the position of the styli is outside the black measuring surface areas of the flags.

Possible cause of error	Solution
The patient's lower jaw was not in reference position when the writing bow was set to the hinge axis point.	Repeat the procedure of adjusting the writing bow. Make sure that the patient keeps their lower jaw closed in centric position.
The holding clamps of the writing bow are not firmly tightened.	Make sure that the holding clamps of the bow are firmly fixated both on the shaft of the clutch and on the side arms. Otherwise, additional stress caused by the mounted styli can cause certain displacements.

The writing bow was not properly adjusted to be free of tension.	Repeat the procedure of adjusting the writing bow. Make sure to keep the system free of tension. The writing bow must not be allowed to move, especially when removing the axis locators.
After mounting, the system indicates that there is no contact between stylus and flag of either side.	
Possible cause of error	Solution
A stylus is not connected.	Plug the stylus into the CADIAX device.
A flag is not connected.	Plug the flag into the CADIAX device.
The right and left measuring sensors are erroneously interchanged.	Insert the plugs correctly.
A stylus is outside the allowed measuring area of the flag due to incorrect mounting.	Repeat mounting the face bow. Make sure that both styli are positioned to stay on the black measuring surface of the flag, also during mediotrusive movements.
A measuring flag is soiled.	Due to the flag being soiled, there is no continuous electrical contact between the tip of the stylus and the measuring surface of the flag. Clean the flag surface according to the provided cleaning instructions.
A stylus is defective.	Contact an authorized GAMMA service center.
A flag is defective.	Contact an authorized GAMMA service center.
The CADIAX device is defective.	Contact an authorized GAMMA service center.
Recorded curves go "backwards" instead of protrusive.	
Possible cause of error	Solution
The right and left measuring sensors are erroneously interchanged.	Insert the plugs correctly.
Recorded curves do not begin in the coordinate origin.	
Possible cause of error	Solution
The position of the styli has changed between setting the reference position and starting the recording.	Repeat setting the reference position. If necessary, repeat this procedure before each recording.
Recorded curves show strong artifacts.	
Possible cause of error	Solution
A measuring flag is soiled.	Due to the flag being soiled, there is no continuous electrical contact between the tip of the stylus and the measuring surface of the flag. Clean the flag surface according to the provided cleaning instructions.
A stylus is outside the allowed measuring area of the flag due to incorrect mounting.	Repeat mounting the face bow. Make sure that both styli are positioned to stay on the black measuring surface of the flag, also during mediotrusive movements.
"Sliding chair effect" caused by surface friction between stylus and flag.	The movement of the stylus across the flag surface can cause a so called "sliding chair effect" to occur. During such, the tip of the stylus shortly lifts up from the flag surface, causing electrical

	contact to be lost. During application, this effect is often accompanied by a squeaking sound during movements. To minimize this effect, apply a smallest possible amount of medical petroleum jelly onto the measuring surface of the flag. Make sure to clean the flag after the recording.
A stylus is defective.	Contact an authorized GAMMA service center.
A flag is defective.	Contact an authorized GAMMA service center.
The CADIAX device is defective.	Contact an authorized GAMMA service center.
Recorded curves are too short on one side.	
Possible cause of error	Solution
A measuring flag is soiled.	Due to the flag being soiled, there is no continuous electrical contact between the tip of the stylus and the measuring surface of the flag. Clean the flag surface according to the provided cleaning instructions.
A stylus is outside the allowed measuring area of the flag due to incorrect mounting.	Repeat mounting the face bow. Make sure that both styli are positioned to stay on the black measuring surface of the flag, also during mediotrusive movements.
The right or left measuring sensors are not connected to the device.	Insert the plugs correctly.
A stylus is defective.	Contact an authorized GAMMA service center.
A flag is defective.	Contact an authorized GAMMA service center.
The CADIAX device is defective.	Contact an authorized GAMMA service center.


6 Background information

In this chapter you will find explanations regarding the terminology and backgrounds of the GAMMA concept.

6.1 The reference coordinate system


In order to aggregate CADIAX condylography data and CADIAS cephalometry tracings, all GAMMA products use the hinge-axis-orbital coordinate system as reference. This allows the reproducible and standardized transfer of recorded data in the analog or virtual articulator.

The basis for the axis-orbital coordinate system is the axis-orbital-plane (AOP), which is defined by the left and right hinge axis points and the left orbital point (lowest point of the orbit's anterior margin). Before a lateral radiography, usually lead markers are placed on these three points to indicate their location in the x-ray image.



The axis-orbital-plane on the cranium and in the articulator. The visible axis point and the orbital point are marked red.


The origin of the coordinate system is located centrally on the hinge axis between right and left condyle. From this origin, the positive X axis extends anterior towards the orbital point, the positive Z axis downwards, and the positive Y axis to the right, from the patient's point of view. The median-sagittal plane coincides with the X and Z axis, the transversal plane with the X and Y axis, and the frontal plane with the Y and Z axis.



The origin of the axis-orbital coordinate system in the articulator.

The movements of the right and left condyle recorded during condylography are visualized in symmetric sub-coordinate systems with identical axis definitions.

When recording movements on the patient, the distance between left and right measurement electronics is naturally greater than the distance between the right and left condylar guidance elements in the articulator. It is for this reason that recordings have to be recalculated to the smaller distance before the software can calculate articulator settings.




The zero points of right and left condyle used when recording condylography.




The reference position set during recording represents the zero point of these sub-coordinate systems. Ideally, the reference position is located on the hinge axis and remains fixed for the duration of the recording. However, due to physiological reasons, most patients are unable to return to the precise starting position after a movement. In these cases it can happen that recordings do not start in the zero point, unless the reference position has been re-fixed before each individual recording.

Besides viewing condylographic recordings in sagittal (X/Z) or transversal (X/Y) projection, the CADIAX applications can also display additional coordinate systems for frontal projection (Y/Z):



The coordinate system with sagittal and transversal view.



Extended coordinate system with sagittal, transversal, and frontal view.

6.2 Exact and anatomic hinge axis


A CADIAX condylography recording can be carried out based on either the exact or the anatomic hinge axis of the temporomandibular joint, depending on the used recording system. In both cases, the measurement takes place in close proximity to the joint itself.

The exact hinge axis (also referred to as kinematic or individual hinge axis) is determined during the recording on the patient. Here, the joint is fixated in retral position and the rotation center of stationary open/close movements is observed. Subsequently, the measurement electronics are adjusted to that position and the reference position is set. For recording on exact hinge axis without limitations, a CADIAX 4 or CADIAX diagnostic device is required.

In contrast to the exact hinge axis, the anatomic hinge axis (also referred to as arbitrary hinge axis) is not ascertained for each individual patient. Rather, its location is given through the empirically determined distance of 10 mm anterior to the ear canal. This position is indicated on the Reference AB face bow that is usually used for anatomic recordings and therefore easy to find and adjust to. Recordings on the anatomical hinge axis can be carried out with the devices CADIAX compact 2 as well as CADIAX 4.

The advantage of recording on the anatomic hinge axis is in its simple and quick application. While the deviation from the exact hinge axis inevitably leads to a distortion of the curve characteristics, its effects on the movements used for articulator programming, namely protrusion and mediotrusion, are tolerable.

However, when recording condylography for the purpose of functional diagnostics and occlusion analysis, the exact determination of the patient's individual hinge axis is imperative. In this use case, hinge axis deviations of just a few millimeters in combination with the rotational component can cause significantly different movement patterns in the tooth area.




The effects of various deviations from the exact hinge axis onto the movements of dental cusps.

6.3 Immediate side shift (ISS)

Some articulator systems divide the transversal Bennett curve into an immediate and a subsequent progressive side shift movement. These two phases of the movement are referred to as "immediate side shift" (ISS) and "progressive side shift" (PSS), respectively. The separation is made at a particular distance measured in X direction, which is referred to as "ISS threshold" (ISS-TH).

In the articulator settings calculated by the software, the ISS describes the maximum transversal displacement up to the specified threshold distance. The PSS however is an angle value equatable to the transversal condylar inclination (TCI), not including the initial displacement.



The ISS threshold (ISS, here: 0.5 mm) separates the Bennett movement in immediate (ISS) and progressive side shift (PSS).

7 Specifications

7.1 Technical data

ADC resolution:	14 bit (CADIAX compact 2, model A) 16 bit (CADIAX compact 2, model B) 16 bit (CADIAX 4, model A)
Measurement data display:	0.01 mm
Distance accuracy:	$\pm 5 \%$
Angle accuracy:	$\pm 1.5^\circ$
Supply voltage:	5 V DC (in compliance with USB specifications)
Current input:	max. 0.5 A (in compliance with USB specifications)
Protection class:	B



7.2 Cleaning

7.2.1 CADIAX device

Make sure that the CADIAX device is disconnected from the computer. Also disconnect the sensors and the foot switch from the device. Remove dust and dirt with a soft cloth.



Do not use liquid, and make sure that no liquid gets inside the device during cleaning. The device must never be cleaned under running water or any other liquid. Do not use stiff brushes or steel wool.



Never clean the device with alcohol, organic solvents, or disinfectants, as these may cause damage to the components and the housing.

7.2.2 Measuring flags

The flag surface must not be damaged (e.g., scratched) or soiled (oil, dust, fingerprints, etc.). Check the condition of the flag surface before every application.

If necessary, the flag surface has to be cleaned. To do so, wipe the measuring area with rubbing alcohol and dapple it dry with a soft, lint-free cloth. Leave it in the open afterwards for it to dry completely. Make sure not to damage the sensitive surface during cleaning.



Never treat or clean the flags with stiff brushes or other abrasive materials.

7.2.3 Measuring styli

Only wipe off the styli with a dry, lint-free cloth. If necessary, clean the tips with rubbing alcohol.



Never treat the sockets and the axes with oil or any other lubricants.

7.2.4 Face bow

Parts of the face bow must only be disinfected by spray or wipe disinfection. Wash off any residues under running water. Only the glabella support may be sterilized in an autoclave (5 minutes at 134 °C; 20 minutes at 120 °C).

7.2.5 Functional occlusal clutch

Disinfect the functional occlusal clutch in a disinfecting tank or use disinfectant spray. Subsequently, clean the registration material off the clutch. Sterilization can be carried out in an autoclave (5 minutes at 134 °C; 20 Minutes at 120 °C).

7.2.6 Occlusal tray clutch

Disinfect the occlusal tray clutch in a disinfecting tank or use disinfectant spray. Subsequently clean the registration material off the clutch. Sterilization can be carried out in an autoclave (5 minutes at 134 °C; 20 Minutes at 120 °C).

7.2.7 Bite fork

Disinfect the bite fork in a disinfecting tank or use disinfectant spray. Subsequently clean off any registration material. If a thermoplastic material has been used, it is advisable to place the bite fork in a refrigerator to make removal of the registration material easier. Sterilization can be carried out in an autoclave (5 minutes at 134 °C; 20 Minutes at 120 °C).

7.2.8 3D joint support

The 3D joint support must only be disinfected by wipe disinfection.



Never treat the joints of the 3D joint support with oil or any other lubricants and do not place the part in an immersion bath.

7.3 Recurrent tests

Like all other electronic devices, also the CADIAX device and its electronic components go through the process of aging. Therefore, the measurement function and security of the device has to be tested at regular intervals. A test interval of one year is recommended.

The security testing has to be carried out in compliance with standard IEC 62353 ("Medical electrical equipment - Recurrent test and test after repair of medical electrical equipment") or the appropriate local derivative.

During the recurrent tests, any differences in potential between the touchable, conductive elements of the connector assembly has to be measured. Any detected difference in potential indicates faulty security technology, which has to be corrected. These measurements are carried out as supplementary to the required standard tests above.


Measurement technology testing are carried out by GAMMA or by a service center authorized by GAMMA.

The measuring flags are subjected to relatively high mechanical stresses, and undergo the results of normal attrition. They should be replaced after approximately 500 examinations.


7.4 Operating conditions

The device is designed for operation exclusively in dry areas.

- Temperature: From 10 °C to 35 °C.



- Relative humidity: From 10% to 85%, non-condensing.




In case of a drastic change in temperature, wait until the device has reached room temperature before beginning operation.

7.5 Storage and transport


If the device is to be transported over long distances, pack it in its original packaging. Store the flags and styli in the appropriate containers as long as they are not in use.

For storage and transport, make sure to comply with the following environmental conditions:


- Temperature: From -10 °C to 60 °C.



- Relative humidity: From 10% to 85%. Store in a dry place!



- Atmospheric pressure: From 700 hPa to 1060 hPa.



7.6 Disposal



After their product life, the CADIAX device and its components have to be disposed of correctly. Doing so, comply with the regulations that are in effect in your country or city, regarding the disposal of electronic equipment.



www.gammadental.com